

PREGLED POSTUPKA #114186

1 PODACI O NARUČIOCU

Naziv naručioca	DOO CENTAR ZA EKOTOKSIKOLOŠKA ISPITIVANJA PODGORICA
PIB	02908433
E-mail	info@ceti.co.me
Telefon	020/658-090
Internet adresa	www.ceti.co.me
Fax	020/658-092
Adresa	Bulevar Šarla de Gola 2
Grad	Podgorica
Poštanski broj	81000

2 OSNOVNI PODACI

Opis predmeta javne nabavke	Elisa kitovi -rezidue veterinarskih lijekova i mikotoksina
Status	U izradi
Vrsta predmeta	Robe
Vrsta postupka	Jednostavna nabavka
Službenik za javne nabavke	Milica Laković
Kontakt	020/ 658-090
Datum objave	04.05.2026. 14:15
Napomena	-

3 FAZE U POSTUPKU

Nema faza u izabranom postupku

4 DODATNE INFORMACIJE

Predmet javne nabavke se nabavlja	po partijama
Posebni oblici javne nabavke	
Okvirni sporazum	Ne
Dinamički sistem nabavki	Ne
Elektronska aukcija	Ne
Elektronski katalog	Ne
Nabavka se sprovodi kao	
Zajednička nabavka	Ne
Centralizovana nabavka	Ne
Nabavka je	
Zelena	Ne
Društveno odgovorna	Ne

5 STAVKE PLANA

Godina	Opis	Vrijednost nabavke	Vrijednost PDV	Okvirni sporazum	Vrijednost OS	Vrijednost PDV OS	Vrsta postupka
2026	DOO CENTAR ZA EKOTOKSIKOLOŠKA ISPITIVANJA PODGORICA Elisa kitovi-rezidue veterinarskih ljekova i mikotokisna 33694000 - Dijagnostička sredstva	10.000,00 EUR	2.100,00 EUR	-	-	-	Jednostavna nabavka

6 USLOVI ZA UČEŠĆE U POSTUPKU I ZAHTJEVI U POGLEDU NAČINA IZVRŠAVANJA PREDMETA NABAVKE

Opis	Tip uslova / zahtjeva	Važi za sve partije
Rok izvršenja ugovora je godinu dana od dana potpisivanja ugovora ili utroška nabavkom predviđenih sredstava, u zavisnosti koji od ova dva predviđena uslova bude ranije ispunjen.	Rok izvršenja ugovora	DA
Mjesto izvršenja ugovora je sjedište naručioca, Podgorica, Buleva Šarla de Gola br. 2	Mjesto izvršenja ugovora	DA
Način plaćanja je virmanski na žiro račun dobavljača.	Način plaćanja	DA
Rok plaćanja je 30 dana od dana dostavljanja fakture za isporučenu robu.	Rok plaćanja	DA
Ponuđač je dužan dostaviti dokaz proizvođača ELISA kitova da je zastupnik za Crnu Goru	Drugi uslovi	DA
Ponuđač je dužan dostaviti izjavu o roku trajanja ELISA kitova	Drugi uslovi	DA
Rok trajanja ELISA kita treba biti minimum godinu dana, od dana isporuke	Drugi uslovi	DA
Proizvođač mora imati implementiran ISO standard 17034	Drugi uslovi	DA
Rok isporuke 10 dana od dana potpisivanja Ugovora	Drugi uslovi	DA
U ponudi obavezno navesti naziv proizvođača kao i kataloški broj proizvoda.	Drugi uslovi	DA
Uz ponudu obavezno dostaviti kopiju strane iz kataloga ili kopiju sertifikata za svaku pojedinačnu stavku koja jasno i nedvosmisleno pokazuje da ponuđena roba potpuno zadaoaljava sve karakteristike koje su tražene i specificirane.	Drugi uslovi	DA
Uz ponudu se dostavlja izjava ponuđača o ispunjenosti uslova utvrđenih zahtjevom i nepostojanju sukoba interesa, potpisana od strane ovlašćenog lica ponuđača, data na Obrascu 2 Pravilnika o načinu sprovođenja jednostavnih nabavki, preuzimanjem na posebnom aplikativnom dijelu elektronskog sistema javnih nabavki.	Obrazac 2	DA

7 KRITERIJUMI ZA IZBOR NAJPOVOLJNIJE PONUDE

Opis	Važi za sve partije
Cijena	DA

8 PREDMET NABAVKE

Ukupna procijenjena vrijednost bez PDV:

10.000,00 EUR

Broj partije	1		
Opis partije	Kontaminanti		
Procijenjena vrijednost bez PDV	5.000,00 EUR		
	Opis predmeta nabavke	Bitne karakteristike predmeta nabavke	Količina
1	Aflatoxin M1 sensitive - ELISA Test Kit	Limit detekcije ≤ 10 (ppt) Plate format: 96	3,00 kom
2	Ohratoxin A- ELISA Test Kit	Limit detekcije ≤ 5 (ppb) Plate format: 96	2,00 kom
3	Aflatoxins total- ELISA Test Kit	Limit detekcije ≤ 1 (ppb) Plate format: 96	3,00 kom

Broj partije		2	
Opis partije		Rezidue veterinarskih lijekova i alergeni	
Procijenjena vrijednost bez PDV		5.000,00 EUR	
	Opis predmeta nabavke	Bitne karakteristike predmeta nabavke	Količina
1	Chloramphenicol FAST ELISA Test Kit (Analyte: Chloramphenicol, Chloramphenicol glucuronide)	Cross- reakcija: Chloramphenicol: 100 % Chloramphenicol-glucuronide: 100 % Thiamphenicol: < 1 % Florphenicol: < 1 % Limiti detekcije (ppb) Urin0.02 Mlijeko0.02 Jaja0.02 Tkivo.....0.02 Med0.02	1,00 kom
2	Penicilini ELISA Test Kit (Analyte: ampicillin, penicillin G, amoxicillin, cloxacillin, oxacillin)	Cross- reakcija: Ampicillin 100 %, Benzylpenicillin (Penicillin G) 100 %, Amoxicillin 85 %, Oxacillin 40 %, Cloxacillin 30 % Limiti detekcije (ppb) Mlijeko0.08 Tkivo.....2.5	1,00 kom

3	Tetracycline- ELISA Test Kit (Analyte: tetracycline, oxytetracycline, chlortetracycline, doxytetracycline)	Cross- reakcija: Tetracycline 100 % 4-epitetracycline 87 % 4-epioxytetracycline 52 % Oxytetracycline 52 % Chlortetracycline 51 % Doxycycline 23 % 4-epichlortetracycline 20 % Limiti detekcije (ppb) Mlijeko.....0.4 Med1.7 Tkivo.....2.9 Jaja4.0	1,00 kom
4	Gluten- ELISA Test Kit (Cocktail if necessary for determination) (Analyte: gliadin)	Limit detekcije ≤ 5 (ppm) Plate format: 96	1,00 kom
5	MALACHITE GREEN TOTAL/CRYSTAL VIOLET ELISA	Antibodi cross-reakcija: Malachite green 100 % Leucomalachite green 100 % Crystal violet 65 % Leucocrystal violet 65 % Karakteristike testa: Matrix LOD (ng/g) Riba i škampi 0.12	1,00 kom

6	COLISTIN ELISA	Antibodi cross-reakcija: Colistin A and B 100 % Karakteristike testa: Matrix - CCβ (ng/ml) Mlijeko - 10 Jaje - 30 Pileće meso - 30 Goveđe meso - 30 Riba - 30	1,00 kom
---	----------------	--	----------