

Univerzitet Crne Gore - Rektorat

Broj : 01/14-227/22

Mjesto i datum : Podgorica, 21.12.2022.g.

Na osnovu člana 143 stav 1 Zakona o javnim nabavkama ("Službeni list CG", broj 74/19), ovlašćeno lice Univerziteta Crne Gore, u postupku javne nabavke opreme za međunarodni projekat "Development of Regional Joint Master Program in Maritime Environmental Protection and Management" (MEP&M) Referentni broj projekta: 619239-EPP-1-2020-1-ME-EPPKA2-CBHE-JP Ugovor o grantu broj: Ref. Ares (2020)7988115-29/12/2020 broj tenderske dokumentacije 01/14 -227/2 od 11.11.2022.g., na prijedlog Komisije za sprovođenje postupka javne nabavke, donosi

ODLUKU
O IZBORU NAJPOVOLJNIJE PONUDE ZA PARTIJE 3,9,14

U postupku javne nabavke po tenderskoj dokumentaciji za nabavku ostale opreme međunarodni projekat "Development of Regional Joint Master Program in Maritime Environmental Protection and Management" (MEP&M) Referentni broj projekta: 619239-EPP-1-2020-1-ME-EPPKA2-CBHE-JP Ugovor o grantu broj: Ref. Ares (2020)7988115-29/12/2020 broj tenderske dokumentacije 01/14 -227/2 od 11.11.2022.g.,

- 1) Ponuda ponuđača **BD MANAGEMENT DOO Vračar Beograd** je najpovoljnija;
- 2) Ponuda ponuđača **Raster office d.o.o.** odbija se kao neispravna;

OBRAZLOŽENJE**1. Podaci o naručiocu**

Naziv naručioca	Univerzitet Crne Gore - Rektorat
Ovlašćeno lice naručioca	rektor prof. dr Vladimir Božović
Kontakt osoba	Simonović Željka
Adresa	Cetinjska broj 2
Grad	Podgorica
Poštanski broj	81000
Identifikacioni broj	02016702

2. Podaci o postupku javne nabavke

Broj tenderske dokumentacije	01/14 -227/2 od 11.11.2022.g izmjene i dopune
------------------------------	---

<i>Vrsta postupka</i>	Otvoreni postupak
<i>Vrsta predmeta</i>	roba
<i>Opis predmeta nabavke</i>	nabavka ostale opreme međunarodni projekat "Development of Regional Joint Master Program in Maritime Environmental Protection and Management" (MEP&M)
<i>Šifra postupka</i>	24875
<i>CPV šifra</i>	39300000 - Razna oprema
<i>Predmet javne nabavke se nabavlja</i>	po partijama

<i>Partija 1¹</i>	Partija 1: Nabavka opreme za <i>distance learning</i> za potrebe Univerziteta Crne Gore (Crna Gora), procijenjena vrijednost 31.000,00 eura bez carine i PDV-a.
<i>Partija 2²</i>	Partija 2: Nabavka opreme za podršku <i>distance learning-a</i> za potrebe Univerziteta Crne Gore (Crna Gora), procijenjena vrijednost 10.000,00 eura bez carine i PDV-a.
<i>Partija 3</i>	Partija 3: Nabavka opreme za <i>podmorsko istraživanje</i> za potrebe Univerziteta Crne Gore (Crna Gora), procijenjena vrijednost 37.000,00 eura bez carine i PDV-a.
<i>Partija 4</i>	Partija 4: Nabavka vazdušnog drona za potrebe Univerziteta Crne Gore (Crna Gora), procijenjena vrijednost 3.000,00 eura bez carine i PDV-a.
<i>Partija 5</i>	Partija 5: Upgrade postojećeg sistema za kontrolu kvaliteta vazduha za potrebe Univerziteta Crne Gore (Crna Gora), procijenjena vrijednost 20.000,00 eura bez carine i PDV-a.
<i>Partija 6</i>	Partija 6: Nabavka stručne literature u elektronskoj formi (e-books) za potrebe Univerziteta Crne Gore (Crna Gora), procijenjena vrijednost 4.000,00 eura bez carine i PDV-a.
<i>Partija 7</i>	Partija 7: Nabavka opreme za <i>distance learning</i> za potrebe Univerziteta u Valoni „Ismail

¹ Ukoliko je primjenjivo

² Ukoliko je primjenjivo

		Ćemali“ (Albanija), procijenjena vrijednost 31.000,00 evra bez carina i PDV-a.
Partija 8		Partija 8: Nabavka opreme za podršku <i>distance learning</i> -a za potrebe Univerziteta u Valoni „Ismail Ćemali“ (Albanija), procijenjena vrijednost 6.500,00 eura bez carine i PDV-a.
Partija 9		Partija 9: Nabavka opreme za <i>podmorsko istraživanje</i> za potrebe Univerziteta u Valoni „Ismail Ćemali“ (Albanija), procijenjena vrijednost 26.500,00 evra bez carina i PDV-a.
Partija 10		Partija 10: Nabavka vazdušnog drona za potrebe Univerziteta u Valoni „Ismail Ćemali“ (Albanija), procijenjena vrijednost 3.000,00 eura bez carine i PDV-a.
Partija 11		Partija 11: Nabavka stručne literature u elektronskoj formi za potrebe Univerziteta u Valoni „Ismail Ćemali“ (Albanija), procijenjena vrijednost 3.000,00 eura bez carine i PDV-a.
Partija 12		Partija 12: Nabavka opreme za <i>distance learning</i> za potrebe Univerziteta u Draču „Aleksander Moisiju“ (Albanija), procijenjena vrijednost 31.000,00 evra bez carina i PDV-a.
Partija 13		Partija 13: Nabavka opreme za podršku <i>distance learning</i> -a za potrebe Univerziteta u Draču „Aleksander Moisiju“ (Albanija), procijenjena vrijednost 6.500,00 eura bez carine i PDV-a.
Partija 14		Partija 14: Nabavka opreme za <i>podmorsko istraživanje</i> za potrebe Univerziteta u Draču „Aleksander Moisiju“ (Albanija), procijenjena vrijednost 26.500,00 evra bez carina i PDV-a.
Partija 15		Partija 15: Nabavka vazdušnog drona za potrebe Univerziteta u Draču „Aleksander Moisiju“ (Albanija), procijenjena vrijednost 3.000,00 eura bez carine i PDV-a.
Partija 16		Partija 16: Nabavka stručne literature u elektronskoj formi za potrebe Univerziteta u

	Draču „Aleksander Moisiu“ (Albanija), procijenjena vrijednost 3.000,00 eura bez carine i PDV-a.
<i>Procijenjena vrijednost javne nabavke po partijama</i>	UKUPNO: 245.000,00 evra

3. Podaci o podnesenim inicijalnim/konačnim ponudama³

Nije primjenjivo.

4. Podaci o podnesenim ponudama koje nijesu otvorene, sa razlozima neotvaranja

Nije primjenjivo.

5. Podaci o kandidatima/ponuđačima koji su isključeni iz postupaka javne nabavke iz razloga navedenih u Obavještenju o isključenju broj _____ od dana _____

Nije primjenjivo.

6. Podaci o ponudama koje su neispravne sa razlozima neispravnosti

Neispravna je ponuda ponuđača :

RASTER OFFICE D.O.O. iz razloga kako slijedi:

1. Uvidom u sadržaj bankarske garancije broj 10-0-10391 od 10.11.2022.g. CKB banka za: Partija 3 :740,00 eura; Partija 9: 530 eur Partija 14: 530,00 eur; Ukupno: 1.800,00 eur, navedeno je: „ istek roka garancije 17.01.2023.g. poslije ovog datuma garancija prestaje da važi bez obzira da li je original vraćen na našu adresu ili ne“.

Članom 97 stav 3 Zakona o javnim nabavkama ("Sl. list CG", br. 74/2019) definisano je trajanje garancije ponude "Trajanje garancije ponude određuje se u skladu sa rokom važenja ponude uključujući i rok za eventualno aktiviranje koji ne može da bude duži od osam dana, od dana isteka roka važenja ponude". S tim u vezi, uslovima predmetnog postupka, između ostalog zahtijevano sledeće: „ ponuđač je dužan da dostavi bezuslovnu i na prvi poziv naplativu garanciju ponude u iznosu od 2% procijenjene vrijednosti partije, kao garanciju ostajanja u obavezi prema ponudi u periodu važenja ponude i 8 dana nakon isteka važenja ponude“, pri čemu je rok važenja ponude 60 dana od dana otvaranja ponuda. Otvaranje je održano 11.11.2022.u 15h.

Članom 89 stav 2 Zakona o upravnom postupku ("Sl. list CG", br. 56/2014, 20/2015, 40/2016 I 37/2017) definisan je način računanja rokova „Kad je rok određen na dane, dan u koji je dostavljanje izvršeno, odnosno dan u koji pada događaj od kojeg treba računati trajanje roka, ne uračunava se u rok, već se za početak roka uzima prvi naredni radni dan“.

³ Ukoliko je primjenjivo

S tim u vezi, Komisija konstatuje da je ponuđač dostavio neispravnu bankarsku garanciju u smislu roka trajanja u kojoj je naveden istek roka 17.01.2023.g, umjesto 18.01.2023.g. kako treba da stoji, a što je 68 dana od dana otvaranja ponuda.

Na osnovu navedenog, ponuda se smatra neispravnom u smislu člana 133 stav 1 tačka 3 Zakona o javnim nabavkama ("Sl. list CG", br. 74/2019).

Takodje, u dijelu tehničke specifikacije, utvrđene su nepravilnosti, kako slijedi:

PARTIJA 3

UNDERWATER EQUIPMENT FOR SEAWATER VISUAL AND QUALITY MONITORING WITH ACCESSORIES		
Underwater Remotely Operated Vehicle – ROV with accessories	Underwater remotely operated vehicle – ROV with next minimum characteristics: <ol style="list-style-type: none"> 1 Niskin water sampler or similar, for integration with the ROV – at least 1 l capacity 	Underwater remotely operated vehicle – ROV with next minimum characteristics: <ol style="list-style-type: none"> 1 U brošuri nema podataka o Niskin water sampler-u s obzirom da ga proizvođač ne nudi u okviru svoje ponude na web sajtu
Sea water quality measurement instrument with accessories	Sea water quality measurement instrument with the following characteristics <ol style="list-style-type: none"> 1. Conforms to the following standards: EN61326-1 Class B, EN IEC 63000 9, 2. Capable of measuring following parameters: pH, Dissolved oxygen, Electrical conductivity, Salinity, Total dissolved solid, Seawater specific gravity, Temperature, Turbidity, Water depth, Oxidation reduction potential, GPS location 3. Auto calibration feature 4. Possibility to set a site name for repeated measurements at the same location(s); at least 20 sites can be stored Sensor head characteristics <ol style="list-style-type: none"> 5. Measurement temperature range - 10°C to 55°C 6. Cable length at least 30m Control unit characteristics: <ol style="list-style-type: none"> 7. Outer dimensions (WxDxH) 115 mm x 66 mm x 335 mm ±5% 	Sea water quality measurement instrument with the following characteristics <p><u>Opšti zaključak da ponuđena tehnička specifikacija senzora ne odgovara onoj prikazanoj u brošuri. Takođe, kontrolna jedinica nije potkrijepljena i odgovarajućom brošurom iz koje bi se vidjelo da li ista odgovara ponuđenoj specifikaciji.</u></p> <ol style="list-style-type: none"> 1. U brošuri nije vidljivo da zadovoljava tražene standarde 2. U brošuri nije vidljivo da ima mogućnost mjerenja Turbidity 3. U brošuri se ne vidi mogućnost autokalibracije 4. U brošuri se ne može vidjeti mogućnost skladištenja i imenovanja 20 pozicija 5. U brošuri je ponuđen temp. opseg - 5°C to 50°C. Ovo nije u skladu sa traženom i ponuđenom specifikacijom 6. U brošuri se ne vidi kolika je moguća dužina kabla 7. U brošuri uopšte nije data kontrolna

	<p>8. Mass maximum 825 g</p> <p>9. Monochromatic graphic LCD with backlight 320x240 mm</p> <p>10. Memory: at least 10000 data items</p> <p>11. USB connection</p> <p>12. IP-67 waterproofing, or better</p> <p>13. GPS unit: Reception method (12 channel parallel); Measurement precision [With PDOP (high precision): 30 m or less (2 drms)] or better</p> <p>14. Ambient temperature range: at least -5°C to 45°C</p> <p>DO sensor:</p> <p>15. Measurement method: Polarographic method</p> <p>Electrical conductivity probe:</p> <p>16. Measurement method: Four-AC-electrode method</p> <p>17. Range: 0 S/m to 10 S/m (0 mS/cm to 100 mS/cm) or better</p> <p>Salinity measurement:</p> <p>18. Measurement method: Electrical conductivity conversion</p> <p>19. Range: at least 0 to 70 parts per thousand</p> <p>20. Resolution: 0.01 part per thousand or better</p> <p>21. Precision: ± 3 part per thousand or better</p> <p>TDS (total dissolved solid) measurement:</p> <p>22. Measurement method: Electrical conductivity conversion</p> <p>23. Conversion coefficient setting</p> <p>24. Range: at least 0 g/L to 100 g/L</p> <p>25. Resolution: 0.1% of full-scale or better</p> <p>26. Repeatability: ± 2 g/L or better</p> <p>27. Precision: ± 5 g/L or better</p> <p>Seawater specific gravity measurement:</p>	<p>jedinica, već se uz mjerni senzor mora nositi prenosni računar. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>8. Isti komentar kao i prethodni</p> <p>9. Isti komentar kao i prethodni</p> <p>10. Isti komentar kao i prethodni</p> <p>11. Isti komentar kao i prethodni</p> <p>12. Isti komentar kao i prethodni</p> <p>13. Isti komentar kao i prethodni</p> <p>14. Isti komentar kao i prethodni</p> <p>15. U brošuri je data drugačija mjerna metoda od tražene i ponuđene specifikacijom. Tražena je polarografska a u brošuri je navedena optička metoda (EPA 1002-8-2009)</p> <p>16. U brošuri uopšte nije prikazan način za mjerenje električne provodljivosti. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>17. U brošuri je ponuđen opseg 5 $\mu\text{S/cm}$ to 100.000 $\mu\text{S/cm}$. Ovo nije u skladu sa traženom i ponuđenom specifikacijom (traženo 0 $\mu\text{S/cm}$ – 100.000 $\mu\text{S/cm}$)</p> <p>18. U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje saliniteta. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>19. Isti komentar kao i prethodni</p> <p>20. Isti komentar kao i prethodni</p> <p>21. Isti komentar kao i prethodni</p> <p>22. U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje TDS-a. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>23. Isti komentar kao i prethodni</p> <p>24. Isti komentar kao i prethodni</p> <p>25. Isti komentar kao i prethodni</p> <p>26. Isti komentar kao i prethodni</p>
--	---	--

	<p>28. Measurement method: Electrical conductivity conversion</p> <p>29. σ_t, σ_0, σ_{15} display</p> <p>30. Range: at least 0 σ_t to 50 σ</p> <p>31. Resolution: 0.1 σ or better</p> <p>32. Precision: $\pm 5 \sigma$ or better</p> <p>Temperature probe:</p> <p>33. Measurement method: Platinum temperature sensor</p> <p>34. Range: at least -10°C to 55°C</p> <p>35. Sensor: Platinum temperature sensor, JIS Class B (0.3+0.005 t) or better</p> <p>Turbidity measurement:</p> <p>36. Measurement method: LED 90° scattering method</p> <p>37. Range: at least 0 NTU to 1000 NTU</p> <p>38. Resolution: 0.01 NTU or better</p> <p>39. Precision: $\pm 5\%$ of read-out or ± 1 NTU (whichever is larger) or better</p> <p>40. Conforms to EN ISO 7027</p> <p>ORP (oxidation reduction potential) measurement:</p> <p>41. Measurement method: Platinum electrode method</p> <p>42. Range: at least -2000 mV to +2000 mV</p> <p>Scope of supply:</p> <p>43. One water quality measurement instrument with the above stated characteristics</p> <p>44. Spare pH sensor – appropriate for the instrument</p> <p>45. pH 4.01 Standard Solution for auto calibration, at least 4l</p> <p>46. Spare ORP sensor – appropriate for the instrument</p> <p>47. Spare DO membrane cap (at least 3 pieces) – appropriate for the instrument</p>	<p>27. Isti komentar kao i prethodni</p> <p>28. U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje specifične težine. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>29. Isti komentar kao i prethodni</p> <p>30. Isti komentar kao i prethodni</p> <p>31. Isti komentar kao i prethodni</p> <p>32. Isti komentar kao i prethodni</p> <p>33. U brošuri je data drugačija mjerna metoda od tražene i ponuđene specifikacijom. Tražena je senzor od platine a ponuđena u brošuri je od žive (EPA 170.1)</p> <p>34. U brošuri je ponuđen temp. opseg -5°C to 50°C. Ovo nije u skladu sa traženom i ponuđenom specifikacijom (treba od -10°C to 55°C ili bolje)</p> <p>35. Isti komentar kao pod 33</p> <p>36. U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje mutnoće morske vode. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>37. Isti komentar kao i prethodni</p> <p>38. Isti komentar kao i prethodni</p> <p>39. Isti komentar kao i prethodni</p> <p>40. Isti komentar kao i prethodni</p> <p>41. U brošuri uopšte nije prikazan način za mjerenje ORP. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>42. U brošuri je ponuđen opseg -1400 mV to +1400 mV. Ovo nije u skladu sa traženom i ponuđenom specifikacijom (-2000 mV to +2000 mV)</p> <p>43. Iz brošure se ne može ustanoviti. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>44. Isti komentar kao i prethodni</p>
--	--	---

	48. Spare Reference Electrode tip with liquid junction – appropriate for the instrument 49. Internal solution for DO 50. pH sensor protect cover – 3 pieces at least 51. DO sensor protect cover – 3 pieces at least 52. Rubber Cap for sensor guard – at least 8 pieces 53. Backpack to carry the instrument 54. Carrying case for the instrument – hard plastic with foam inside to protect the instrument, or better	45. Isti komentar kao i prethodni 46. Isti komentar kao i prethodni 47. Isti komentar kao i prethodni 48. Isti komentar kao i prethodni 49. Isti komentar kao i prethodni 50. Isti komentar kao i prethodni 51. Isti komentar kao i prethodni 52. Isti komentar kao i prethodni 53. Isti komentar kao i prethodni 54. Isti komentar kao i prethodni
Underwater device for sound recording with accessories	Underwater device for sound recording with accessories has to meet the following minimal characteristics: Scope of delivery: 55. Appropriate recording device capable of independent recording of the hydrophone signals. Memory at least 512 GB. Working on batteries and external power source.	Underwater device for sound recording with accessories has to meet the following minimal characteristics: 55. U brošuri uopšte nije prikazan uređaj za snimanje signala. Ovo nije u skladu sa traženom i ponudenom specifikacijom

PARTIJA 9

UNDERWATER EQUIPMENT FOR SEAWATER VISUAL AND QUALITY MONITORING WITH ACCESSORIES		
Underwater Remotely Operated Vehicle – ROV with accessories	Underwater remotely operated vehicle – ROV with next minimum characteristics: 2 Niskin water sampler or similar, for integration with the ROV – at least 1 l capacity	Underwater remotely operated vehicle – ROV with next minimum characteristics: 2 U brošuri nema podataka o Niskin water sampler-u s obzirom da ga proizvođač ne nudi u okviru svoje ponude na web sajtu
Sea water quality measurement instrument with accessories	Sea water quality measurement instrument with the following characteristics	Sea water quality measurement instrument with the following characteristics <u>Opšti zaključak da ponuđena tehnička specifikacija senzora ne odgovara onoj</u>

	<p>56. Conforms to the following standards: EN61326-1 Class B, EN IEC 63000 9,</p> <p>57. Capable of measuring following parameters: pH, Dissolved oxygen, Electrical conductivity, Salinity, Total dissolved solid, Seawater specific gravity, Temperature, Turbidity, Water depth, Oxidation reduction potential, GPS location</p> <p>58. Auto calibration feature</p> <p>59. Possibility to set a site name for repeated measurements at the same location(s); at least 20 sites can be stored</p> <p>Sensor head characteristics</p> <p>60. Measurement temperature range - 10°C to 55°C</p> <p>61. Cable length at least 30m</p> <p>Control unit characteristics:</p> <p>62. Outer dimensions (WxDxH) 115 mm x 66 mm x 335 mm ±5%</p> <p>63. Mass maximum 825 g</p> <p>64. Monochromatic graphic LCD with backlight 320x240 mm</p> <p>65. Memory: at least 10000 data items</p> <p>66. USB connection</p> <p>67. IP-67 waterproofing, or better</p> <p>68. GPS unit: Reception method (12 channel parallel); Measurement precision [With PDOP (high precision): 30 m or less (2 drms)] or better</p> <p>69. Ambient temperature range: at least -5°C to 45°C</p> <p>DO sensor:</p> <p>70. Measurement method: Polarographic method</p> <p>Electrical conductivity probe:</p> <p>71. Measurement method: Four-AC-electrode method</p>	<p><u>prikazanoj u brošuri. Takođe, kontrolna jedinica nije potkrijepljena i odgovarajućom brošurom iz koje bi se vidjelo da li ista odgovara ponuđenoj specifikaciji.</u></p> <p>56. U brošuri nije vidljivo da zadovoljava tražene standarde</p> <p>57. U brošuri nije vidljivo da ima mogućnost mjerenja Turbidity</p> <p>58. U brošuri se ne vidi mogućnost autokalibracije</p> <p>59. U brošuri se ne može vidjeti mogućnost skladištenja i imenovanja 20 pozicija</p> <p>60. U brošuri je ponuđen temp. opseg - 5°C to 50°C. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>61. U brošuri se ne vidi kolika je moguća dužina kabla</p> <p>62. U brošuri uopšte nije data kontrolna jedinica, već se uz mjerni senzor mora nositi prenosni računar. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>63. Isti komentar kao i prethodni</p> <p>64. Isti komentar kao i prethodni</p> <p>65. Isti komentar kao i prethodni</p> <p>66. Isti komentar kao i prethodni</p> <p>67. Isti komentar kao i prethodni</p> <p>68. Isti komentar kao i prethodni</p> <p>69. Isti komentar kao i prethodni</p> <p>70. U brošuri je data drugačija mjerna metoda od tražene i ponuđene specifikacijom. Tražena je polarografska a u brošuri je navedena optička metoda (EPA 1002-8-2009)</p> <p>71. U brošuri uopšte nije prikazan način za mjerenje električne provodljivosti. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>72. U brošuri je ponuđen opseg 5 µS/cm to 100.000 µS/cm. Ovo nije u skladu</p>
--	--	--

72. Range: 0 S/m to 10 S/m (0 mS/cm to 100 mS/cm) or better

Salinity measurement:

73. Measurement method: Electrical conductivity conversion

74. Range: at least 0 to 70 parts per thousand

75. Resolution: 0.01 part per thousand or better

76. Precision: ± 3 part per thousand or better

TDS (total dissolved solid) measurement:

77. Measurement method: Electrical conductivity conversion

78. Conversion coefficient setting

79. Range: at least 0 g/L to 100 g/L

80. Resolution: 0.1% of full-scale or better

81. Repeatability: ± 2 g/L or better

82. Precision: ± 5 g/L or better

Seawater specific gravity measurement:

83. Measurement method: Electrical conductivity conversion

84. σ_t , σ_0 , σ_{15} display

85. Range: at least 0 σ_t to 50 σ

86. Resolution: 0.1 σ or better

87. Precision: ± 5 σ or better

Temperature probe:

88. Measurement method: Platinum temperature sensor

89. Range: at least -10°C to 55°C

90. Sensor: Platinum temperature sensor, JIS Class B ($0.3+0.005 |t|$) or better

Turbidity measurement:

91. Measurement method: LED 90° scattering method

sa traženom i ponuđenom specifikacijom (traženo $0 \mu\text{S/cm}$ – $100.000 \mu\text{S/cm}$)

73. U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje saliniteta. Ovo nije u skladu sa traženom i ponuđenom specifikacijom

74. Isti komentar kao i prethodni

75. Isti komentar kao i prethodni

76. Isti komentar kao i prethodni

77. U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje TDS-a. Ovo nije u skladu sa traženom i ponuđenom specifikacijom

78. Isti komentar kao i prethodni

79. Isti komentar kao i prethodni

80. Isti komentar kao i prethodni

81. Isti komentar kao i prethodni

82. Isti komentar kao i prethodni

83. U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje specifične težine. Ovo nije u skladu sa traženom i ponuđenom specifikacijom

84. Isti komentar kao i prethodni

85. Isti komentar kao i prethodni

86. Isti komentar kao i prethodni

87. Isti komentar kao i prethodni

88. U brošuri je data drugačija mjerna metoda od tražene i ponuđene specifikacijom. Tražena je senzor od platine a ponuđena u brošuri je od žive (EPA 170.1)

89. U brošuri je ponuđen temp. opseg -5°C to 50°C . Ovo nije u skladu sa traženom i ponuđenom specifikacijom (treba od -10°C to 55°C ili bolje)

90. Isti komentar kao pod 33

91. U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje

	<p>92. Range: at least 0 NTU to 1000 NTU</p> <p>93. Resolution: 0.01 NTU or better</p> <p>94. Precision: $\pm 5\%$ of read-out or ± 1 NTU (whichever is larger) or better</p> <p>95. Conforms to EN ISO 7027</p> <p>ORP (oxidation reduction potential) measurement:</p> <p>96. Measurement method: Platinum electrode method</p> <p>97. Range: at least -2000 mV to +2000 mV</p> <p>Scope of supply:</p> <p>98. One water quality measurement instrument with the above stated characteristics</p> <p>99. Spare pH sensor – appropriate for the instrument</p> <p>100. pH 4.01 Standard Solution for auto calibration, at least 4l</p> <p>101. Spare ORP sensor – appropriate for the instrument</p> <p>102. Spare DO membrane cap (at least 3 pieces) – appropriate for the instrument</p> <p>103. Spare Reference Electrode tip with liquid junction – appropriate for the instrument</p> <p>104. Internal solution for DO</p> <p>105. pH sensor protect cover – 3 pieces at least</p> <p>106. DO sensor protect cover – 3 pieces at least</p> <p>107. Rubber Cap for sensor guard – at least 8 pieces</p> <p>108. Backpack to carry the instrument</p> <p>109. Carrying case for the instrument – hard plastic with foam inside to protect the instrument, or better</p>	<p>mutnoće morske vode. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>92. Isti komentar kao i prethodni</p> <p>93. Isti komentar kao i prethodni</p> <p>94. Isti komentar kao i prethodni</p> <p>95. Isti komentar kao i prethodni</p> <p>96. U brošuri uopšte nije prikazan način za mjerenje ORP. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>97. U brošuri je ponuđen opseg -1400 mV to +1400 mV. Ovo nije u skladu sa traženom i ponuđenom specifikacijom (-2000 mV to +2000 mV)</p> <p>98. Iz brošure se ne može ustanoviti. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>99. Isti komentar kao i prethodni</p> <p>100. Isti komentar kao i prethodni</p> <p>101. Isti komentar kao i prethodni</p> <p>102. Isti komentar kao i prethodni</p> <p>103. Isti komentar kao i prethodni</p> <p>104. Isti komentar kao i prethodni</p> <p>105. Isti komentar kao i prethodni</p> <p>106. Isti komentar kao i prethodni</p> <p>107. Isti komentar kao i prethodni</p> <p>108. Isti komentar kao i prethodni</p> <p>109. Isti komentar kao i prethodni</p>
--	--	---

UNDERWATER EQUIPMENT FOR SEAWATER VISUAL AND QUALITY MONITORING WITH ACCESSORIES

Underwater Remotely Operated Vehicle – ROV with accessories	Underwater remotely operated vehicle – ROV with next minimum characteristics: 3 Niskin water sampler or similar, for integration with the ROV – at least 1 l capacity	Underwater remotely operated vehicle – ROV with next minimum characteristics: 3 U brošuri nema podataka o Niskin water sampler-u s obzirom da ga proizvođač ne nudi u okviru svoje ponude na web sajtu
Sea water quality measurement instrument with accessories	Sea water quality measurement instrument with the following characteristics 110.Conforms to the following standards: EN61326-1 Class B, EN IEC 63000 9, 111.Capable of measuring following parameters: pH, Dissolved oxygen, Electrical conductivity, Salinity, Total dissolved solid, Seawater specific gravity, Temperature, Turbidity, Water depth, Oxidation reduction potential, GPS location 112.Auto calibration feature 113.Possibility to set a site name for repeated measurements at the same location(s); at least 20 sites can be stored Sensor head characteristics 114.Measurement temperature range - 10°C to 55°C 115.Cable length at least 30m Control unit characteristics: 116.Outer dimensions (WxDxH) 115 mm x 66 mm x 335 mm ±5% 117.Mass maximum 825 g 118.Monochromatic graphic LCD with backlight 320x240 mm 119.Memory: at least 10000 data items 120.USB connection 121.IP-67 waterproofing, or better	Sea water quality measurement instrument with the following characteristics <u>Opšti zaključak da ponuđena tehnička specifikacija senzora ne odgovara onoj prikazanoj u brošuri. Takođe, kontrolna jedinica nije potkrijepljena i odgovarajućom brošurom iz koje bi se vidjelo da li ista odgovara ponuđenoj specifikaciji.</u> 110.U brošuri nije vidljivo da zadovoljava tražene standarde 111.U brošuri nije vidljivo da ima mogućnost mjerenja Turbidity 112.U brošuri se ne vidi mogućnost autokalibracije 113.U brošuri se ne može vidjeti mogućnost skladištenja i imenovanja 20 pozicija 114.U brošuri je ponuđen temp. opseg - 5°C to 50°C. Ovo nije u skladu sa traženom i ponuđenom specifikacijom 115.U brošuri se ne vidi kolika je moguća dužina kabla 116.U brošuri uopšte nije data kontrolna jedinica, već se uz mjerni senzor mora nositi prenosni računar. Ovo nije u skladu sa traženom i ponuđenom specifikacijom 117.Isti komentar kao i prethodni 118.Isti komentar kao i prethodni 119.Isti komentar kao i prethodni 120.Isti komentar kao i prethodni

122. GPS unit: Reception method (12 channel parallel); Measurement precision [With PDOP (high precision): 30 m or less (2 drms)] or better

123. Ambient temperature range: at least -5°C to 45°C

DO sensor:

124. Measurement method: Polarographic method

Electrical conductivity probe:

125. Measurement method: Four-AC-electrode method

126. Range: 0 S/m to 10 S/m (0 mS/cm to 100 mS/cm) or better

Salinity measurement:

127. Measurement method: Electrical conductivity conversion

128. Range: at least 0 to 70 parts per thousand

129. Resolution: 0.01 part per thousand or better

130. Precision: ± 3 part per thousand or better

TDS (total dissolved solid) measurement:

131. Measurement method: Electrical conductivity conversion

132. Conversion coefficient setting

133. Range: at least 0 g/L to 100 g/L

134. Resolution: 0.1% of full-scale or better

135. Repeatability: ± 2 g/L or better

136. Precision: ± 5 g/L or better

Seawater specific gravity measurement:

137. Measurement method: Electrical conductivity conversion

138. σ_t , σ_0 , σ_{15} display

139. Range: at least 0 σ_t to 50 σ

140. Resolution: 0.1 σ or better

141. Precision: ± 5 σ or better

121. Isti komentar kao i prethodni

122. Isti komentar kao i prethodni

123. Isti komentar kao i prethodni

124. U brošuri je data drugačija mjerna metoda od tražene i ponuđene specifikacijom. Tražena je polarografska a u brošuri je navedena optička metoda (EPA 1002-8-2009)

125. U brošuri uopšte nije prikazan način za mjerenje električne provodljivosti. Ovo nije u skladu sa traženom i ponuđenom specifikacijom

126. U brošuri je ponuđen opseg 5 $\mu\text{S/cm}$ to 100.000 $\mu\text{S/cm}$. Ovo nije u skladu sa traženom i ponuđenom specifikacijom (traženo 0 $\mu\text{S/cm}$ – 100.000 $\mu\text{S/cm}$)

127. U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje saliniteta. Ovo nije u skladu sa traženom i ponuđenom specifikacijom

128. Isti komentar kao i prethodni

129. Isti komentar kao i prethodni

130. Isti komentar kao i prethodni

131. U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje TDS-a. Ovo nije u skladu sa traženom i ponuđenom specifikacijom

132. Isti komentar kao i prethodni

133. Isti komentar kao i prethodni

134. Isti komentar kao i prethodni

135. Isti komentar kao i prethodni

136. Isti komentar kao i prethodni

137. U brošuri uopšte nije prikazana

specifikacija i metoda za mjerenje specifične težine. Ovo nije u skladu sa traženom i ponuđenom specifikacijom

138. Isti komentar kao i prethodni

139. Isti komentar kao i prethodni

	<p>Temperature probe:</p> <p>142.Measurement method: Platinum temperature sensor</p> <p>143.Range: at least -10°C to 55°C</p> <p>144.Sensor: Platinum temperature sensor, JIS Class B (0.3+0.005 t) or better</p> <p>Turbidity measurement:</p> <p>145.Measurement method: LED 90° scattering method</p> <p>146.Range: at least 0 NTU to 1000 NTU</p> <p>147.Resolution: 0.01 NTU or better</p> <p>148.Precision: ±5% of read-out or ±1 NTU (whichever is larger) or better</p> <p>149.Conforms to EN ISO 7027</p> <p>ORP (oxidation reduction potential) measurement:</p> <p>150.Measurement method: Platinum electrode method</p> <p>151.Range: at least -2000 mV to +2000 mV</p> <p>Scope of supply:</p> <p>152.One water quality measurement instrument with the above stated characteristics</p> <p>153.Spare pH sensor – appropriate for the instrument</p> <p>154.pH 4.01 Standard Solution for auto calibration, at least 4l</p> <p>155.Spare ORP sensor – appropriate for the instrument</p> <p>156.Spare DO membrane cap (at least 3 pieces) – appropriate for the instrument</p> <p>157.Spare Reference Electrode tip with liquid junction – appropriate for the instrument</p> <p>158.Internal solution for DO</p>	<p>140.Isti komentar kao i prethodni</p> <p>141.Isti komentar kao i prethodni</p> <p>142.U brošuri je data drugačija mjerna metoda od tražene i ponuđene specifikacijom. Tražena je senzor od platine a ponuđena u brošuri je od žive (EPA 170.1)</p> <p>143.U brošuri je ponuđen temp. opseg -5°C to 50°C. Ovo nije u skladu sa traženom i ponuđenom specifikacijom (treba od -10°C to 55°C ili bolje)</p> <p>144.Isti komentar kao pod 33</p> <p>145.U brošuri uopšte nije prikazana specifikacija i metoda za mjerenje mutnoće morske vode. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>146.Isti komentar kao i prethodni</p> <p>147.Isti komentar kao i prethodni</p> <p>148.Isti komentar kao i prethodni</p> <p>149.Isti komentar kao i prethodni</p> <p>150.U brošuri uopšte nije prikazan način za mjerenje ORP. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>151.U brošuri je ponuđen opseg -1400 mV to +1400 mV. Ovo nije u skladu sa traženom i ponuđenom specifikacijom (-2000 mV to +2000 mV)</p> <p>152.Iz brošure se ne može ustanoviti. Ovo nije u skladu sa traženom i ponuđenom specifikacijom</p> <p>153.Isti komentar kao i prethodni</p> <p>154.Isti komentar kao i prethodni</p> <p>155.Isti komentar kao i prethodni</p> <p>156.Isti komentar kao i prethodni</p> <p>157.Isti komentar kao i prethodni</p> <p>158.Isti komentar kao i prethodni</p> <p>159.Isti komentar kao i prethodni</p> <p>160.Isti komentar kao i prethodni</p> <p>161.Isti komentar kao i prethodni</p>
--	--	--

	159.pH sensor protect cover – 3 pieces at least 160.DO sensor protect cover – 3 pieces at least 161.Rubber Cap for sensor guard – at least 8 pieces 162.Backpack to carry the instrument 163.Carrying case for the instrument – hard plastic with foam inside to protect the instrument, or better	162.Isti komentar kao i prethodni 163.Isti komentar kao i prethodni
--	--	--

Na osnovu navedenog, ponuda se smatra neispravnom u smislu člana 133 stav 1 tačka 4 Zakona o javnim nabavkama ("Sl. list CG", br. 74/2019).

7. Podaci o ispravnim ponudama

Komisija za sprovođenje postupka javne nabavke je u postupku pregleda I ocjene ponude ponuđača BD MANAGEMENT DOO BEOGRAD VRAČAR utvrdila sledeće:

- ponuda ponuđača BD MANAGEMENT DOO BEOGRAD VRAČAR sačinjena na način predviđen tenderskom dokumentacijom;
- ponudjač je dostavio Izjavu privrednog subjekta, koja je sačinjena na obrascu datom u „Pravilniku o obrascu izjave privrednog subjekta" "Službeni list Crne Gore", br. 071/20 od 16. jula 2020. godine. Ponuđač je tačno i nedvosmisleno popunio Izjavu privrednog subjekta u skladu sa zahtjevom iz tenderske dokumentacije, čime je dokazao ispunjenost uslova za učešće u postupku javne nabavke.
- U Izjavi privrednog subjekta, u dijelu IV Sukob interesa, navedeno je da na strani ovog privrednog subjekta NE postoji razlog koji predstavlja sukob interesa iz člana 41 Zakona o javnim nabavkama.

Stručno lice konstatuje da sve stavke navedene u finansijskom dijelu ponude u cjelosti odgovaraju opisu, bitnim karakteristikama i količinama iz specifikacije. Takođe, privredni subjekat je dostavio tehničku dokumentaciju, katalog robe koja je predmet nabavke. Uvidom u dostavljeni katalog utvrđeno je da ponuđena oprema odgovara zahtijevanoj. Dostavljena je autorizacija proizvođača opreme:

1.Letter of Authorization for distribution Date: 14/10/2022 Project Number : 619239-EPP-1-2020-1-ME-EPPKA2-CBHE-JP Tender number : 01/14-227/2 To: The University of Montenegro WHEREAS We HORIBA UK Limited, who are official manufacturers of water quality testing instruments under the product line names of LAQUA and U-50, having factories at 31, Miyanonishi-cho, Kisshoin, Minami-ku, Kyoto 601-8306 Japan do hereby authorize BD Management doo, having its offices at Kralja Milutina 9, 11000 Belgrade, Serbia with VAT number 109449342 and company registration number 21184497 to submit a bid the purpose of which is to provide the following goods, U-54G(30M) / 3200323688 / U-54 for measurement of pH, ORP, Dissolved Oxygen, Conductivity, Salinity, TDS, Seawater specific gravity and Temperature with 30 meter cable, supplied with 1 x pH 4 (500ml), 1 x pH reference solution (330 - 250ml), 1 x DO solution set (50ml internal sol, sandpaper, syringe), cleaning brush, calibration cup, Back pack, Strap, 4 x

Alkaline batteries (LR14), Silicon grease and instruction manual manufactured by us and to subsequently negotiate and sign the contract.

Horiba UK LIMITED -pismo ovlaštenja za distribuciju - prevedeno od strane sudskog prevodioca za engleski jezik.

2.Letter of Authorization for Distribution Date: October 17, 2022 Project Number: 619239-EPP-1-2020-1-ME-EPPKA2-CBHE-JP Tender Number: 01/14-227/2 To: The University of Montenegro We Blue Robotics, who are official manufacturers of the BlueROV2 remotely operated underwater vehicle, having a facility at 2740 California Street Torrance, CA, USA, do hereby authorize BD Management doo, having its offices at Kralja Milutina 9, 11000 Belgrade, Serbia with VAT number 109449342 and company registration number 21184497 to submit a bid the purpose of which is to provide the following goods, BlueROV2 with options, accessories, and spare parts, manufactured by us and to subsequently negotiate and sign the contract. Signed: Name: Kristel Hall Title: Customer Success Lead Blue Robotics Dated on 17th day of October

Pismo ovlaštenja za distribuciju BLUE ROBOTICS, prevedeno od strane sudskog prevodioca za engleski jezik.

Potvrda specifikacije BlueRobotics partija 3

Potvrda specifikacije BlueRobotics partije 9 i 14

Potvrda specifikacije partije 3, 9 i 14 Horiba

Potvrda specifikacije DolphinEar partija 3

- Ponuđač je neposrednom predajom na arhivi dostavio tehničku dokumentaciju, katalog robe koja je predmet nabavke, u skladu sa uslovima predviđenim predmetnom nabavkom i to:
- Letter of technical specification confirmation Blue robotics confirmation, lot 3
- Letter of technical specification confirmation Blue ROBOTICS confirmation, lot 9, 14
- Letter of compliance HORIBA za partije 3, 9, 14
- HORIBA multi-parameter water quality meters specification;
- BLUEROV 2 technical specification;
- General oceanics partija 3 details, description
- DolphinEar hydrophones DE200 Series specifications
- TASCAM partija 3 specifications;
- Ponuđač je neposrednom predajom na arhivi dostavio garanciju ponude uz naznaku na CEJNU da je garancija dostavljena neposredno. Garancija je izdata od strane BANCA INTESA BROJ 10860336 od 08.11.2022.G. za Partija 3, 9 i 14 ukupno: 1.800,00 eur sa rokom važenja 18.01.2023.g., čime je ispunjen zahtjev naručioca u pogledu iznosa visine garancije od 2% procijenjene vrijednosti nabavke, načina dostavljanja i roka važenja u periodu važenja ponude i 8 dana nakon isteka roka važenja ponude.
- ponuđač je u finansijskom dijelu ponude ponudio specifikaciju u skladu sa tehničkom specifikacijom iz tenderske dokumentacije (opisan je predmet nabavke, navedene bitne karakteristike, jedinice mjere i količine).
- Ponuđač je u pogledu načina izvršavanja predmeta nabavke u ponudi naveo uslove tražene u uslovima za učešće u predmetnom postupku.

U postupku ocjene ispravnosti ponude Komisija za sprovođenje postupka javne nabavke je u skladu sa članom 112 Zakona o javnim nabavkama Sl. List Crne Gore 074/19 izvršila provjeru tačnosti podataka navedenih u izjavi.

Od ponuđača⁴ **BD MANAGEMENT DOO BEOGRAD VRAČAR** koji je podnio ekonomski najpovoljniju ponudu zahtijevano je da u roku od 8 dana od dana prijema zahtjeva dostavi originale ili ovjerene kopije dokaza predviđenih tenderskom dokumentacijom. Datum upućivanja zahtjeva je 18.11.2022.g. broj akta 01/14-227/11.

Ponuđač **BD MANAGEMENT DOO BEOGRAD VRAČAR** je u predviđenom roku dana 24.11.2022g. dostavio sljedeće dokaze:

Uvjerenje broj 388/2022 od 29.07.2022. da na osnovu podataka koji se vode u službenoj evidenciji osnovnih i viših sudova uključujući Posebno Odjeljenje za organizovani kriminal i Odjeljenje za ratne zločine Višeg suda u Beogradu i posebna odjeljenja za suzbijanje korupcije Višeg suda u Beogradu, Višeg suda u Nišu, Višeg suda u Kraljevu i Višeg suda u Novom Sadu da pravno lice **BD MANAGEMENT D.O.O. Vračar Beograd** nije osuđivano za krivična dela iz nadležnosti ovih sudova- original;

Uvjerenje Republika Srbija Ministarstvo unutrašnjih poslova, Sektor za analitiku, telekomunikacione i informacione tehnologije, Odjeljenje za analitiku, telekom, informacione tehnologije u Novom Sadu, Odsjek za analitiku i policijske evidencije u PU **SUBOTICA** broj 05.17.13-235-96228/22 od 09.11.2022. da **Nikolić Dušan** prema podacima iz kaznene evidencije ovog Ministarstva nije osuđivan-ovjerena kopija.

Uvjerenje Republike Srbije Ministarstvo finansija poreska uprava –odjeljenje za poresko računovodstvo broj 252-437-08-08456/2022 od 17.10.2022.g. da je obveznik **BD MANAGEMENT D.O.O. Vračar Beograd** PIB 109449342 prema računovodstvenoj evidenciji uplatio dospelje obaveze javnih prihoda evidentirane u poreskom računovodstvu zaključno sa 17.10.2022.g.-original.

Uvjerenje Grada Beograda Republika Srbija, Gradska uprava Grada Beograda, Sekretarijat za javne prihode, Sektor javnih prihoda za područja gradskih opština, odjeljenje Palilula broj XXVI-01-437-180/2022 od 25.10.2022. godine kojim se potvrđuje da se na dan 25.10.2022.godine je obveznik **BD MANAGEMENT D.O.O. Vračar Beograd** PIB 109449342 nalazi u evidenciji po osnovu izvornih lokalnih javnih prihoda na teritoriji gradske opštine Vračar i izmirene su dospelje obaveze na osnovu istih zaključno sa 25.10.2022. -original.

Izvod o registraciji privrednog subjekta izdat dana 20.07.2022. godine u 10:26 h od strane Agencije za privredne registre Republike Srbije kojim se potvrđuje aktivan status ponuđača „ **BD MANAGEMENT D.O.O. Vračar Beograd** matični broj 21184497, PIB 109449342-original.

Letter of Authorization for distribution Date: 14/10/2022 Project Number : 619239-EPP-1-2020-1-ME-EPPKA2-CBHE-JP Tender number : 01/14-227/2 To: The University of Montenegro WHEREAS We **HORIBA UK Limited**, who are official manufacturers of water quality testing instruments under the product line names of **LAQUA** and **U-50**, having factories at 31, Miyanonishi-cho, Kisshoin, Minami-ku, Kyoto 601-8306 Japan do hereby authorize **BD Management doo**, having its offices at **Kralja Milutina 9, 11000 Belgrade, Serbia** with VAT number 109449342 and company registration number 21184497 to submit a bid the purpose of which is to provide the following goods, **U-54G(30M) / 3200323688 / U-54** for measurement of pH, ORP, Dissolved Oxygen, Conductivity, Salinity, TDS, Seawater specific gravity and Temperature with 30 meter cable, supplied with 1 x pH

⁴ Ponoviti za svakog sljedećeg ponuđača ukoliko prvorangirani ne zadovoljava

4 (500ml), 1 x pH reference solution (330 - 250ml), 1 x DO solution set (50ml internal sol, sandpaper, syringe), cleaning brush, calibration cup, Back pack, Strap, 4 x Alkaline batteries (LR14), Silicon grease and instruction manual manufactured by us and to subsequently negotiate and sign the contract.

Horiba UK LIMITED -pismo ovlaštenja za distribuciju - prevedeno od strane sudskog prevodioca za engleski jezik- Igor Prlja.

Letter of Authorization for Distribution Date: October 17, 2022 Project Number: 619239-EPP-1-2020-1-ME-EPPKA2-CBHE-JP Tender Number: 01/14-227/2 To: The University of Montenegro We Blue Robotics, who are official manufacturers of the BlueROV2 remotely operated underwater vehicle, having a facility at 2740 California Street Torrance, CA, USA, do hereby authorize BD Management doo, having its offices at Kralja Milutina 9, 11000 Belgrade, Serbia with VAT number 109449342 and company registration number 21184497 to submit a bid the purpose of which is to provide the following goods, BlueROV2 with options, accessories, and spare parts, manufactured by us and to subsequently negotiate and sign the contract. Signed: Name: Kristel Hall Title: Customer Success Lead Blue Robotics Dated on 17th day of October.

Pismo ovlaštenja za distribuciju BLUE ROBOTICS, prevedeno od strane sudskog prevodioca za engleski jezik Igor PRLJA.

Na osnovu izloženog, ponuda ponuđača „ BD MANAGEMENT D.O.O. Vračar Beograd je u skladu sa članom 132 Zakona o javnim nabavkama Sl. List 074/19, ocijenjena ispravnom.

Metodologija vrednovanja primijenjena u postupku je odnos cijene i kvaliteta koje se vrši na osnovu sljedećih parametara:

1. Parametar: Cijena (C)maksimalan broj bodova 90
2. Parametar: Kvalitet (K)maksimalan broj bodova 10

Ukupan broj bodova = (C) + (K)

Parametar cijena: Ponuđaču koji ponudi najnižu cijenu dodjeljuje se maksimalan broj bodova, dok ostali ponuđači dobijaju proporcionalni broj bodova u odnosu na najnižu ponuđenu cijenu, prema formuli:

Broj bodova = $\frac{\text{Najniža ponuđena cijena} \times 90}{\text{Ponuđena cijena}}$

Parametar kvalitet: Vrednovanje ponuda po osnovu parametra kvalitet odnosi se na rok isporuke. Ponuđač sa najkraćim ponuđenim rokom isporuke dobija maksimalan broj bodova. Ostale prispjele ponude biće bodovane prema formuli:

Broj bodova = $\frac{\text{najkraći ponuđeni rok za isporuku robe} \times 10}{\text{ponuđeni rok za isporuku robe}}$

Napomena: Ponuđeni rok isporuke ne može biti duži od 60 dana od dana potpisivanja ugovora. Na osnovu dostavljenih izvještaja svih članova komisije u skladu sa metodologijom utvrđen je prosjek dodijeljenog broja bodova po partijama:

Partija 3

r.b.	Ponudáč (naziv i sjedište)	Parametar 1	Parametar 2 ⁵	Ukupan broj bodova
1.	BD MANAGEMENT DOO BEOGRAD VRAČAR	90	10	100

Partija 9

r.b.	Ponudáč (naziv i sjedište)	Parametar 1	Parametar 2 ⁶	Ukupan broj bodova
1.	BD MANAGEMENT DOO BEOGRAD VRAČAR	90	10	100

Partija 14

r.b.	Ponudáč (naziv i sjedište)	Parametar 1	Parametar 2 ⁷	Ukupan broj bodova
1.	BD MANAGEMENT DOO BEOGRAD VRAČAR	90	10	100

8. Podaci o najpovoljnijoj ponudi

BD MANAGEMENT DOO BEOGRAD VRAČAR

Partija 3: Nabavka opreme za podmorsko istraživanje za potrebe Univerziteta Crne Gore (Crna Gora), procijenjena vrijednost 37.000,00 eura bez carine i PDV-a.

Ponudjena cijena: 37.000 EUR

⁵ Dodati onoliko koliko je parametara predviđeno tenderskom dokumentacijom i za svaki upisati broj bodova

⁶ Dodati onoliko koliko je parametara predviđeno tenderskom dokumentacijom i za svaki upisati broj bodova

⁷ Dodati onoliko koliko je parametara predviđeno tenderskom dokumentacijom i za svaki upisati broj bodova

Rok isporuke : 60 dana od dana potpisivanja ugovora.

Garantni rok navedeno u bitnim karakteristikama: 12 mjeseci.

Mjesto izvršenja ugovora: Mesto izvršenja u skladu sa zahtevima naručioca

Rok plaćanja je 30 dana od dana ispostavljanja uredne fakture sa otpremnicom i zapisnikom o izvršenoj kvalitativnoj i kvantitativnoj isporuci

Način plaćanja: Avans u iznosu 50%, ostatak virmanski po izvršenoj isporuci

Partija 9: Nabavka opreme za podmorsko istraživanje za potrebe Univerziteta u Valoni „Ismail Ćemali“ (Albanija), procijenjena vrijednost 26.500,00 evra bez carina i PDV-a.

Ponudjena cijena: 26.500,00 EUR

Rok isporuke : 60 dana od dana potpisivanja ugovora.

Garantni rok navedeno u bitnim karakteristikama: 12 mjeseci.

Mjesto izvršenja ugovora: Mesto izvršenja u skladu sa zahtevima naručioca

Rok plaćanja je 30 dana od dana ispostavljanja uredne fakture sa otpremnicom i zapisnikom o izvršenoj kvalitativnoj i kvantitativnoj isporuci.

Način plaćanja: Avans u iznosu 50%, ostatak virmanski po izvršenoj isporuci

Partija 14: Nabavka opreme za podmorsko istraživanje za potrebe Univerziteta u Draču „Aleksander Moisiu“ (Albanija), procijenjena vrijednost 26.500,00 evra bez carina i PDV-a.

Ponudjena cijena: 26.500,00 EUR

Rok isporuke : 60 dana od dana potpisivanja ugovora.

Garantni rok navedeno u bitnim karakteristikama: 12 mjeseci.

Mjesto izvršenja ugovora: Mesto izvršenja u skladu sa zahtevima naručioca

Rok plaćanja je 30 dana od dana ispostavljanja uredne fakture sa otpremnicom i zapisnikom o izvršenoj kvalitativnoj i kvantitativnoj isporuci.

Način plaćanja: Avans u iznosu 50%, ostatak virmanski po izvršenoj isporuci

9. Podaci o najpovoljnijoj/im ponudi/ama u slučaju zaključivanja okvirnog sporazuma sa jednim ili više ponuđača⁸

/

⁸ Ukoliko je primjenjivo

10. Uputstvo o pravnoj zaštiti

Kandidat/ponuđač može izjaviti žalbu protiv ove odluke Komisiji za zaštitu prava u roku od 10 dana od dana objavljivanja odluke.

Žalba se podnosi naručiocu preko ESJN. Žalba koja nije podnesena na naprijed predviđeni način biće odbijena kao nedozvoljena.

Podnosilac žalbe je dužan da uz žalbu priloži dokaz o uplati naknade za vođenje postupka u iznosu od 1% od procijenjene vrijednosti javne nabavke, a najviše 20.000,00 eura, na žiro račun Komisije za zaštitu prava broj 530-20240-15 kod NLB Montenegro banke A.D.

Iznos uplate iz prethodnog stava je iznos koji je prispio na račun Komisije za zaštitu prava.

Ukoliko je predmet nabavke podijeljen po partijama, a žalba se odnosi samo na određenu/e partiju/e, naknada se plaća u iznosu 1% od procijenjene vrijednosti javne nabavke te /tih partije/a.

Instrukcije za plaćanje naknade za vođenje postupka od strane žalilaca iz inostranstva nalaze se na internet stranici Komisije za zaštitu prava <http://www.kontrola-nabavki.me/>.

OVLAŠĆENO LICE NARUČIOCA

rektor prof. dr Vladimir Božović



(ime i prezime)
(potpis)⁹
Božović

⁹ Ukoliko naručilac koristi sertifikovani elektronski potpis, pečat nije potreban